



# 730S I.V. House UltraDressing<sup>®</sup>

1 Small Hand Size IV Site Protector - 730 Series 1 Puppy Sticker and 1 Hand Measurement Chart (In Box)

U.S. Patents Nos 6.526.981 / 8.006.699 / 8.640.707 Canadian Patent No. 2.415.102 Additional Patents Pending All Rights Reserved • Made in USA

Video Directions and Info: www.ivhouse.com/how

Manufactured for I.V. House, Inc. 2072 Concourse Dr St. Louis. MO 63146 USA IVinfo@ivhouse.com (E-mail)

(F

800-530-0400 (toll free) 314-453-9200 (phone) 314-453-9576 (fax)



Wellkang Ltd, Enterprise Hub, NW Business Complex, 1 Beraghmore Rd, Derry, BT48 8SE, Northern Ireland








SINGLE USE MRI SAFE











## 730S I.V. House UltraDressing® Directions

### Single Use Only-Fits Either Hand

Using hand measurement chart, measure patient's hand knuckle to knuckle between 1st & 4th finger to find the size.

**Right hand**: Slip thumb through thumbhole closest to dome. **Left hand**: Slip thumb through thumbhole farthest from dome.

- Following IV insertion, tape the loop of tubing small enough to fit entirely inside the inner curve of the dome.
- Position the plastic dome over the looped tubing and catheter hub (round edge of dome towards fingertips). Wrap fabric snugly around palm, securing with Velcro<sup>®</sup>.
- Release the Velcro tab to inspect the IV site according to your nursing practice guidelines.
- Change the I.V. House UltraDressing for each new IV insertion or if soiled.

### Puppy Sticker

Remove the crack and peel backing from the puppy sticker and apply the sticker to the dome.

Tell the patient that the puppy likes to sleep on the "house" and to leave her alone.

#### Note:

#### It is important that the looped tubing is covered by the dome

to prevent accidental snagging of the loop and catheter hub. This also prevents bending and occlusion. Covering both the hub and the looped tubing virtually eliminates catheter manipulation.





For Detailed Directions www.ivhouse.com/how